



November 26, 2019

**ADDENDUUM ACKNOWLEDGEMENT FORM
ITB 2019-2020-02 ADDENDUM #2**

Proposal No: ITB 2019-2020-02
Proposal Title: Radiography Equipment
Revised Proposal Due Date: December 11, 2019 2:00 p.m. CST
Questions Due Date: December 4, 2019 4:30 p.m. CST
Revised Proposal Opening: December 11, 2019 2:00 p.m. CST

PLEASE BE ADVISED THAT THE FOLLOWING CHANGES ARE APPLICABLE TO THE ORIGINAL SPECIFICATIONS OF THE ABOVE-REFERENCED ITB:

This addendum includes the following:

RESPONSE TO WITTEN QUESTIONS RECEIVED OR ADDITIONAL BID DOCUMENTS:

- 1. Question:** Item #4 on the summary sheet ask us to include installation cost including Overhead Crane, Unistrut, Electrical and Demonstration. The Overhead Crane is part of the equipment and is included but the Unistrut and Electrical will need to be in place before we can install the equipment. Do you want us to provide the Unistrut and Electrical specifications and you have a contractor install it?

Response: Vendor will be required to install electrical, unistrut bracing, everything required for complete install and operation. If that means vendor needs to hire an electrical contractor to do the work, then they need to hire an electrical contractor as part of their package.

THIS ADDENDUM NOW BECOMES A PART OF THE ORIGINAL ITB.

THE ADDENDUM ACKNOWLEDGMENT FORM SHALL BE SIGNED BY AN AUTHORIZED COMPANY REPRESENTATIVE, DATED AND RETURNED WITH THE RESPONSE.

COMPANY NAME: _____

AUTHORIZED SIGNATURE: _____ DATE: _____